

VME based DAQ system for ICAL RPC detector stack at TIFR

M.Bhuyan¹, V.B.Chandratre², S.Dasgupta¹, S.D.Kalmani¹, S.M.Lahamge¹, N.K.Mondal¹,
P.Nagaraj¹, S.K.Rao¹, A.Redij¹, D.Samuel¹, M.N.Saraf¹, B.Satyanarayana¹,
R.R.Shinde¹ and S.S.Upadhy¹

1: Department of High Energy Physics, Tata Institute of Fundamental Research, Mumbai, 400005, India.

2: Electronics Division, Bhabha Atomic Research Centre, Mumbai, 400085, India.

Email: sudeshnadasgupta@tifr.res.in

A 12 layer 1m x1m glass RPC stack is operational and collecting cosmic ray data for the last three years at TIFR. A VME based DAQ system with various graphical interfaces for data acquisition and monitoring has been developed. Both hardware and software aspects of this system will be discussed in this paper.