

VME based DAQ system for INO RPC stack

M.Bhuyan^a, V.B.Chandratre^b, S.Dasgupta^a, V.M.Datar^c, S.D.Kalmani^a, S.M.Lahamge^a,
N.K.Mondal^a, P.Nagaraj^a, S.K.Rao^a, A.Redij^a, D.Samuel^{a*}, M.N.Saraf^a,
B.Satyanarayana^a, R.R.Shinde^a and S.S.Upadhy^a

^aDepartment of High Energy Physics, Tata Institute of Fundamental Research, Mumbai 400005, INDIA

^bElectronics Division, Bhabha Atomic Research Centre, Mumbai 400085, INDIA

^cNuclear Physics Division, Bhabha Atomic Research Centre, Mumbai 400085, INDIA

A 12 layer 1m x1m glass RPC stack is operational and collecting cosmic ray data for the last two years at TIFR. A VME based DAQ system with various graphical interfaces for data acquisition and monitoring has been developed. Both hardware and software aspects of this system will be discussed in this paper.

* Corresponding author
Email address: samuel@tifr.res.in (Deepak Samuel)